

Date: July 13, 2009

Dear Sir or Madam:

Please be advised that the TAXUS<sup>®</sup> Liberté<sup>®</sup> Long Paclitaxel-Eluting Stent Systems (Monorail<sup>®</sup> and OTW) have undergone testing to detect the presence of Natural Rubber Latex. Based on the "Below Detectable" result produced during this testing, neither the finished TAXUS Liberté Long Paclitaxel-Eluting Stent Systems (Monorail and OTW) product nor their packaging are believed to contain any detectable Natural Rubber Latex.

If you have any further inquiries or if we may be of any further assistance, please do not hesitate to contact our Customer Service Center at 888-272-1001.

Sincerely,



Adam Heroux  
Manager Biocompatibility Services  
Boston Scientific Corporation

Attachment

## Product Inquiry

Finished Good / Product I.D. Number	Product Description	Natural Rubber Latex Content
H7493893638270	TAXUS Liberte Long (MR) 2.75mm x 38mm	BD
H7493893638300	TAXUS Liberte Long (MR) 3.00mm x 38mm	BD
H7493893638350	TAXUS Liberte Long (MR) 3.50mm x 38mm	BD
H7493893638400	TAXUS Liberte Long (MR) 4.00mm x 38mm	BD
H7493893738270	TAXUS Liberte Long (OTW) 2.75mm x 38mm	BD
H7493893738300	TAXUS Liberte Long (OTW) 3.00mm x 38mm	BD
H7493893738350	TAXUS Liberte Long (OTW) 3.50mm x 38mm	BD
H7493893738400	TAXUS Liberte Long (OTW) 4.00mm x 38mm	BD
	End of Inquiry	

Key:	
<b>Present</b>	Natural Rubber Latex is believed to be present in the product or its packaging based on a review of the bill of materials for that product and its packaging and/or testing of the finished product and its packaging. As required by regulatory authorities, each such product or packaging bears a caution regarding the presence of Natural Rubber Latex to alert the user and/or patients with a known sensitivity.
<b>Not in BOM or BD</b>	The finished product is believed to contain no Natural Rubber Latex based on: a review of the bill of materials for that product and its packaging; and/or a “Below Detection” result from testing of the finished product and its packaging.