

Date : 21-May-2009

Dear Sir or Madam:

Please be advised that the TAXUS[®] Liberté[®] Atom[™] Paclitaxel-Eluting Stent System (Monorail[®] and OTW) has undergone testing to detect the presence of Natural Rubber Latex. Based on the “Below Detectable” result produced during this testing, neither the finished TAXUS[®] Liberté[®] Atom[™] Paclitaxel-Eluting Stent System (Monorail[®] and OTW) product nor its packaging is believed to contain any detectable Natural Rubber Latex.

If you have any further inquiries or if we may be of any further assistance, please do not hesitate to contact our Customer Service Center at 888-272-1001.

Sincerely,



Adam Heroux
Manager Biocompatibility Services
Boston Scientific Corporation

Attachment

Product Inquiry

Finished Good / Product I.D. Number	Product Description	Natural Rubber Latex Content
H749389368220	TAXUS Liberte Atom (MR) 2.25mm x 8mm	BD
H7493893612220	TAXUS Liberte Atom (MR) 2.25mm x 12mm	BD
H7493893616220	TAXUS Liberte Atom (MR) 2.25mm x 16mm	BD
H7493893620220	TAXUS Liberte Atom (MR) 2.25mm x 20mm	BD
H7493893624220	TAXUS Liberte Atom (MR) 2.25mm x 24mm	BD
H7493893628220	TAXUS Liberte Atom (MR) 2.25mm x 28mm	BD
H7493893632220	TAXUS Liberte Atom (MR) 2.25mm x 32mm	BD
H749389378220	TAXUS Liberte Atom(OTW) 2.25mm x 8mm	BD
H7493893712220	TAXUS Liberte Atom (OTW) 2.25mm x 12mm	BD
H7493893716220	TAXUS Liberte Atom (OTW) 2.25mm x 16mm	BD
H7493893724220	TAXUS Liberte Atom (OTW) 2.25mm x 20mm	BD
H7493893724220	TAXUS Liberte Atom (OTW) 2.25mm x 24mm	BD
H7493893728220	TAXUS Liberte Atom (OTW) 2.25mm x 28mm	BD
H7493893732220	TAXUS Liberte Atom (OTW) 2.25mm x 32mm	BD

Key:	
Present	Natural Rubber Latex is believed to be present in the product or its packaging based on a review of the bill of materials for that product and its packaging and/or testing of the finished product and its packaging. As required by regulatory authorities, each such product or packaging bears a caution regarding the presence of Natural Rubber Latex to alert the user and/or patients with a known sensitivity.
Not in BOM or BD	The finished product is believed to contain no Natural Rubber Latex based on: a review of the bill of materials for that product and its packaging; and/or a “Below Detection” result from testing of the finished product and its packaging.